

# Sunstone Scientific LLC



## NIST Traceable Calibration for Scattering Sensor

### 1. DESCRIPTION

The scattering sensor is calibrated to provide accurate and reproducible measurements of the volume scattering function (VSF) with an angular weighting defined by a centroid angle  $\bar{\theta}$  and full-width-half-maximum (FWHM) spread  $\Delta\theta$ . Details of the calibration methodology can be found in Twardowski et al. (2012) and Sullivan et al. (2013).

### 2. PHYSICAL DATA

Sensor serial #:	BBFL2b 200
Angular weighting centroid angle $\bar{\theta}$ (assumed):	124°
Angular weighting spread $\Delta\theta$ (assumed):	41°
Spectral weighting centroid wavelength $\bar{\lambda}$ (from manufacturer):	660nm
NIST-traceable beads (cat #: 3100A) lot #:	170408
NIST-traceable beads, certified diameter:	100 ± 3nm

## Certification

$$\beta(\theta_c) [\text{m}^{-1}\text{sr}^{-1}] = \text{Scale Factor} \times (\text{Raw Counts} - \text{Dark Counts})$$

This certifies the following specifications for this scattering sensor:

Scale Factor for 660nm:	3.81E-06 m <sup>-1</sup> sr <sup>-1</sup> counts <sup>-1</sup>
Dark Counts for 660nm:	48.1 counts
Instrument Resolution for 660nm:	1.0 counts
Instrument Resolution for 660nm:	3.81E-06 m <sup>-1</sup> sr <sup>-1</sup>
Approximate worst case accuracy for 660nm (Considering the assumptions above*):	±2.0%

Certification date: 3/2/2018

*Alberto Tonizzo*

Alberto Tonizzo, PhD, *Metrology Director*

\*Based on estimated accuracy of theoretical computation, accuracy of experimental calibration result, and assuming exact geometric consistency between sensors of the same model.