



## NIST Traceable Calibration for Scattering Sensor

### 1. DESCRIPTION

The scattering sensor is calibrated to provide accurate and reproducible measurements of the volume scattering function (VSF) with an angular weighting defined by a centroid angle  $\bar{\theta}$  and full-width-half-maximum (FWHM) spread  $\Delta\theta$ . Details of the calibration methodology can be found in Twardowski et al. (2012) and Sullivan et al. (2013).

### 2. PHYSICAL DATA

Sensor serial #:	BB9-132
Angular weighting centroid angle $\bar{\theta}$ (assumed):	124°
Angular weighting spread $\Delta\theta$ (assumed):	41°
Spectral weighting centroid wavelength $\bar{\lambda}$ (from manufacturer):	407, 439, 485, 507, 527, 594, 651, 715, 878nm
NIST-traceable beads (cat #: 3100A) lot #:	199582
NIST-traceable beads, certified diameter:	100 ± 6nm

## Certification

$$\beta(\theta_c) [\text{m}^{-1}\text{sr}^{-1}] = \text{Scale Factor} \times (\text{Raw Counts} - \text{Dark Counts})$$

This certifies the following specifications for this scattering sensor:

	407nm	439nm	485nm
Scale Factor [ $\text{m}^{-1} \text{sr}^{-1} \text{counts}^{-1}$ ]:	1.45E-05	1.45E-05	1.27E-05
Dark Counts [counts]:	43.8	54.0	49.2
Instrument Resolution [counts]:	0.4	0.3	0.6
Instrument Resolution [ $\text{m}^{-1} \text{sr}^{-1}$ ]:	5.82E-06	4.32E-06	7.05E-06
Approximate best case accuracy (Considering the assumptions above*):	±2.3%	±2.3%	±2.2%
Water temperature during characterization	23.0°C		
Certification date:	11/19/2018		
Expiration date:	11/19/2019		

*Alberto Tonizzo*

Alberto Tonizzo, PhD, *Metrology Director*



## Certification

$$\beta(\theta_c) [\text{m}^{-1} \text{sr}^{-1}] = \text{Scale Factor} \times (\text{Raw Counts} - \text{Dark Counts})$$

This certifies the following specifications for this scattering sensor:

	507nm	527nm	594nm
Scale Factor [ $\text{m}^{-1} \text{sr}^{-1} \text{counts}^{-1}$ ]:	9.74E-06	8.72E-06	5.11E-06
Dark Counts [counts]:	62.5	55.3	54.0
Instrument Resolution [counts]:	0.5	1.1	1.0
Instrument Resolution [ $\text{m}^{-1} \text{sr}^{-1}$ ]:	4.46E-06	9.64E-06	5.15E-06
Approximate best case accuracy (Considering the assumptions above*):	±2.2%	±2.3%	±2.2%

Water temperature during characterization 23.0°C

Certification date: 11/19/2018

Expiration date: 11/19/2019

*Alberto Tonizzo*

Alberto Tonizzo, PhD, *Metrology Director*



## Certification

$$\beta(\theta_c) [\text{m}^{-1} \text{sr}^{-1}] = \text{Scale Factor} \times (\text{Raw Counts} - \text{Dark Counts})$$

This certifies the following specifications for this scattering sensor:

	651nm	715nm	878nm
Scale Factor [ $\text{m}^{-1} \text{sr}^{-1} \text{counts}^{-1}$ ]:	3.95E-06	3.16E-06	2.55E-06
Dark Counts [counts]:	51.8	54.2	52.5
Instrument Resolution [counts]:	0.6	0.4	0.6
Instrument Resolution [ $\text{m}^{-1} \text{sr}^{-1}$ ]:	2.23E-06	1.36E-06	1.49E-06
Approximate best case accuracy (Considering the assumptions above*):	±2.2%	±2.2%	±2.1%

Water temperature during characterization 23.0°C

Certification date: 11/19/2018

Expiration date: 11/19/2019

*Alberto Tonizzo*

Alberto Tonizzo, PhD, *Metrology Director*

\*Based on estimated accuracy of theoretical computation, accuracy of experimental calibration result, and assumptions of geometric consistency between sensors of the same model.