

CALIBRATION REPORT
Portable Radiation Package

SERIAL NUMBER 02

DataSet Configuration: 0010
Document date: February 8, 2001

Configuration File:

PRP CALIBRATION INFORMATION FILE -- created: 2001-02-08 (039) 06:46:52
LAST EDIT: new file
CALIBRATION NAME: 0010
DATA POINT OF CONTACT: R. M. Reynolds
POC ADDRESS: 490D; Brookhaven National lab; Upton NY 11973; USA
POC EMAIL: reynolds@bnl.gov
PRP SERIAL NUMBER: 02
FILE NAME: INFO_02_0010.txt
DOCUMENTS: Cal02_0010.pdf
HEAD SERIAL NUMBER: 432
HEAD CALIBRATION ID: 9904
DATALOGGER SERIAL NUMBER: dl00_4
DATALOGGER CALIBRATION ID: 0010
PSP SERIAL NUMBER: 32385F3
PSP CALIBRATION ID: orig
PIR SERIAL NUMBER: 32387F3
PIR THERMOPILE CALIBRATION ID: orig
PIR TEMPERATURE CALIBRATION ID: 0010
COMMENTS: Revised file structure and matlab



Bldg 490d, Upton NY 11973 — 631-344-7836 — reynolds@bnl.gov

PRP INFO FILE

SETUP FOR PROCESSING PRP CALIBRATIONS: 08-Feb-2001 06:51:34
PRP S/N: 02, Calibration identifier: 0010
Configuration file: hd:instruments:prp:prpcal:prp:02:0010:INFO_02_0010.txt
PRP CALIBRATION
PRP CALIBRATION INFORMATION FILE -- created: 2001-02-08 (039) 06:46:52
LAST EDIT: new file
CALIBRATION NAME: 0010
DATA POINT OF CONTACT: R. M. Reynolds
POC ADDRESS: 490D; Brookhaven National lab; Upton NY 11973; USA
POC EMAIL: reynolds@bnl.gov
PRP SERIAL NUMBER: 02
FILE NAME: INFO_02_0010.txt
DOCUMENTS: Cal02_0010.pdf
HEAD SERIAL NUMBER: 432
HEAD CALIBRATION ID: 9904
DATALOGGER SERIAL NUMBER: dl00_4
DATALOGGER CALIBRATION ID: 0010
PSP SERIAL NUMBER: 32385F3
PSP CALIBRATION ID: orig
PIR SERIAL NUMBER: 32387F3
PIR THERMOPILE CALIBRATION ID: orig
PIR TEMPERATURE CALIBRATION ID: 0010
COMMENTS: Revised file structure and matlab

CALIBRATION INFO FOR HEAD 432:

This file : 432.CAL
Data valid from date : 10/05/1998
MFRSR system owner : PNL / S/O 240
YESDAS system password: Langley!
Supervisor password : Irradiance!
System Datalogger ID : \$0000 (Hex), 0 (Dec)
Instrument Head ID : \$091E (Hex), 2334 (Dec)
Instrument Head S/N : 432

DATALOGGER CALIBRATION: ProcLoggerCal (version 101) Run date: 08-Feb-2001 06:52:59
DATALOGGER S/N DL2000_4
CALIBRATION DATE: 20000507
TECHNICIAN: EDWARDS
VOLTAGE REFERENCE: VOLT-A-VIDER 103264
USE PRECISION VREF CIRCUIT

CALIBRATION INFO FOR HEAD 432:

This file : 432.CAL
Data valid from date : 10/05/1998
MFRSR system owner : PNL / S/O 240
YESDAS system password: Langley!
Supervisor password : Irradiance!
System Datalogger ID : \$0000 (Hex), 0 (Dec)
Instrument Head ID : \$091E (Hex), 2334 (Dec)
Instrument Head S/N : 432

```

-----
DATALOGGER CALIBRATION: ProcLoggerCal (version 101)  Run date: 08-Feb-2001 06:52:59
DATALOGGER S/N DL2000_4
CALIBRATION DATE: 20000507
TECHNICIAN: EDWARDS
VOLTAGE REFERENCE: VOLT-A-VIDER 103264
  USE PRECISION VREF CIRCUIT

```

```

-----
PSP CALIBRATION: S/N 32385F3
  Factory calibration: 1998-02-12
8.92
  8.920

```

```

-----
PIR CALIBRATION - S/N: 32387F3
  Factory calibration: 1998-01-13
3.75
  3.750

```

```

-----
PIR THERMISTOR CALIBRATION -- PART 2

```

```
Process times from file lacie:instruments:prp:prpcal:pir:32387F3:32387F3_Tcal_0010.dat
```

```
Process times from pirSN = 32387F3
```

```
SAVE THE RESULTS TO FILE lacie:instruments:prp:prpcal:pir:32387F3:32387F3_Tcal_0010.mat
```

```
save lacie:instruments:prp:prpcal:pir:32387F3:32387F3_Tcal_0010.mat readme beta_case beta_dome rcav rcs
```

```
  beta_case and beta_dome are Steinhart-Hart coefficients
```

```
  tcase = steinhart(hart(beta_case,r) where r = thermistor resistance in ohms.
```

```
  tdome = steinhart(hart(beta_dome ...
```

```
  rcav and rcstd are the test resistances and the std dev for each test point.
```

```
  rdav and rdstd are ditto for the dome
```

```
  tav and tstd are the mean temperature and std dev for each test point as
```

```
  computed from the SBE834 temperature probe.
```

```
  tc_ysi and td_ysi are computed from the published YSI calibration table.
```

```
  (See function 'YSI44006.m'.)
```

```
  casefit and domefit are cubic fitted corrections to the YSI computed temperatures.
```

```
  If tcaseysi is a vector of case temperatures computed from the YSI44006.m function,
```

```
  then tcase = tcase + polyval(casefit,tcaseysi);
```

```
  Ditto for the dome.
```

```

-----
COMPUTE PIR THERMISTOR COEFFICIENTS

```

```
Use calibrated coefficients for PIR case thermistor
```

```
Use calibrated coefficients for PIR doome thermistor
```

ZENITH ANGLE ERROR PLOTS

Head S/N: 432

Cal date: 10/05/1998

Now: 08-Feb-2001 06:52:15




Figure 1: Zenith Angle Error




Figure 2: Zenith Angle Error

HEAD 432 TOA IRRADIANCES BASED ON ASTRONOMICAL SOLAR SPECTRUM

WAVELENGTH (nm)			IRRADIANCE (W/m ² /nm)		
LOWER	CENTER	UPPER	LOWER	MEAN	UPPER
406,	414,	422,	1.651,	1.738,	1.825
489,	497,	504,	1.846,	1.944,	2.041
605,	614,	624,	1.624,	1.709,	1.795
664,	672,	680,	1.450,	1.526,	1.602
857,	869,	878,	0.903,	0.950,	0.998
927,	938,	947,	0.783,	0.824,	0.865







Figure 3: Head and Logger combined gains



$$1/(T+T_0) = p_1 a^3 + p_2 a^2 + p_3 a + p_4$$

$$a = \ln(\text{mvadc}), T_0 = 273.15$$

Case: max err = 0.009 C

$$p_1 = 2.2006e-06, p_2 = -2.3219e-05$$

$$p_3 = 3.6328e-04, p_4 = 1.9694e-03$$

Dome: max err = 0.007

$$p_1 = 2.0049e-06, p_2 = -2.0262e-05$$

$$p_3 = 3.4660e-04, p_4 = 1.9984e-03$$

Figure 4: Head and Logger combined gains





Figure 5: PIR Temperature calibration results.



TEST PLUG:

$R_{case} =$ _____ $T_{44031} =$ _____ Logger Temperature = _____

$R_{dome} =$ _____ $T_{44031} =$ _____ Logger Temperature = _____

COMPUTE VREF FROM R_REF AND R_THERM

PRP SN: 02 Cal ID: 0010

CASE - Rref = 499400.0

R_cal	v_T	V_REF (computed)
6000	59.1	4979.0
8000	78.5	4978.9
10000	97.7	4976.8
12000	116.8	4977.6
15000	145.1	4977.0
20000	191.7	4978.2
25000	237.3	4977.0
30000	282.0	4975.7
35000	325.8	4974.5

DOME - Rref = 499400.0

R_cal	v_T	V_REF (computed)
6000	59.1	4980.7
8000	78.6	4983.3
10000	97.8	4984.0
12000	117.0	4984.4
15000	145.4	4985.6
20000	192.0	4987.0
25000	237.7	4986.6
30000	282.5	4985.9
35000	326.5	4985.6

(File: hd:instruments:prp:prpcal:prp:02:0010:TcalVref_0010.dat)

```
% CALIBRATION FILE FOR PRPRX DATA COLLECTION SOFTWARE
% PSP CALIBRATION, PSP SN: 32385F3
3.8966      -0.080714
% PIR CALIBRATION, PIR SN: 32387F3
1.276  5.878
% TCASE FIT
2.20065e-06  -2.32188e-05  0.000363276  0.00196936
% TDOME FIT
2.00488e-06  -2.02618e-05  0.000346603  0.00199836
% K COEFFICIENT
4.0
% SIGMA
5.67e-8
% EPSILON
0.98
% BATTERY
0.030820  0.0
```

(File: hd:instruments:prp:prpcal:prp:02:0010:prprx_02_0010.txt)

LOGGER CALIBRATION FILE

```

%PRP02 -- ID = 0010          25 97.42 0.51 98.75 -1.33
%MFRSR: 432                 50 196.17 0.38 197.50 -1.33
%PSP: 32385F3, COEFF: 8.92  75 294.94 0.42 296.25 -1.31
%PIR: 32387F3, COEFF: 3.75  100 394.00 0.59 395.00 -1.00
%PREAMP: 1                  150 591.67 0.48 592.50 -0.83
DATALOGGER S/N DL2000_4     200 789.29 0.47 790.00 -0.71
CALIBRATION DATE: 20000507  250 987.00 0.56 987.50 -0.50
TECHNICIAN: EDWARDS        300 1184.84 0.37 1185.00 -0.16
VOLTAGE REFERENCE: VOLT-A-VIDER 103264 350 1382.33 0.49 1382.50 -0.17
                                400 1580.14 0.57 1580.00 0.14
                                450 1777.95 0.23 1777.50 0.45
                                500 1975.42 0.51 1975.00 0.42
CHANNEL 1                    550 2173.17 0.51 2172.50 0.67
0 0.00 0.00                 600 2370.74 0.45 2370.00 0.74
10 48.09 0.52 49.30 -1.21  650 2568.63 0.50 2567.50 1.13
50 245.49 0.52 246.50 -1.01  700 2766.39 0.50 2765.00 1.39
75 368.68 0.48 369.75 -1.07  750 2964.11 0.46 2962.50 1.61
100 492.25 0.52 493.00 -0.75  800 3161.47 0.51 3160.00 1.47
150 738.84 0.60 739.50 -0.66  850 3359.60 0.50 3357.50 2.10
200 985.79 0.42 986.00 -0.21  900 3556.89 0.57 3555.00 1.89
250 1232.75 0.44 1232.50 0.25
300 1479.05 0.52 1479.00 0.05
350 1725.94 0.24 1725.50 0.44
400 1972.61 0.50 1972.00 0.61
450 2219.16 0.60 2218.50 0.66
500 2465.94 0.25 2465.00 0.94
550 2712.56 0.51 2711.50 1.06
600 2959.37 0.50 2958.00 1.37
650 3206.00 0.33 3204.50 1.50
700 3452.68 0.48 3451.00 1.68
750 3699.35 0.59 3697.50 1.85
CHANNEL 2                    CHANNEL 4
0 0.00 0.00 0.00 0.00      0 0.00 0.00 0.00 0.00
1 8.65 0.55 9.89 -1.24     1 1.50 0.51 2.96 -1.46
10 97.55 0.51 98.90 -1.35  5 13.11 0.46 14.80 -1.69
50 493.56 0.51 494.50 -0.94  10 27.79 0.42 29.60 -1.81
75 740.67 0.52 741.75 -1.08  25 72.45 0.69 74.00 -1.55
100 987.94 0.43 989.00 -1.06  50 146.46 0.59 148.00 -1.54
150 1482.90 0.31 1483.50 -0.60  75 220.16 0.50 222.00 -1.84
200 1977.63 0.50 1978.00 -0.37  100 293.84 0.50 296.00 -2.16
250 2472.32 0.48 2472.50 -0.18  200 589.42 0.51 592.00 -2.58
300 2966.89 0.32 2967.00 -0.11  300 884.84 0.37 888.00 -3.16
350 3461.67 0.49 3461.50 0.17  400 1180.26 0.45 1184.00 -3.74
400 3956.47 0.51 3956.00 0.47  500 1475.78 0.43 1480.00 -4.22
CHANNEL 3                    600 1771.47 0.51 1776.00 -4.53
0 0.00 0.00 0.00 0.00     700 2067.16 0.50 2072.00 -4.84
1 2.55 0.51 3.95 -1.40     800 2362.39 0.61 2368.00 -5.61
10 38.17 0.59 39.50 -1.33  900 2658.16 0.60 2664.00 -5.84
                                1000 2953.58 0.51 2960.00 -6.42
                                1100 3249.30 0.57 3256.00 -6.70
                                1200 3544.47 0.51 3552.00 -7.53
                                1300 3840.32 0.67 3848.00 -7.68
CHANNEL 5
0 0.00 0.00 0.00 0.00
1 1.40 0.50 2.50 -1.10
5 11.10 0.55 12.50 -1.40
10 23.74 0.45 25.00 -1.26

```

25	61.35	0.59	62.50	-1.15	1000	2946.58	0.51	2960.00	-13.42
50	124.00	0.47	125.00	-1.00	1100	3241.26	0.45	3256.00	-14.74
75	186.58	0.51	187.50	-0.92	1200	3536.16	0.37	3552.00	-15.84
100	249.16	0.50	250.00	-0.84	1300	3830.68	0.58	3848.00	-17.32
200	499.89	0.32	500.00	-0.11	PSP				
300	750.26	0.56	750.00	0.26	-5	-143.86	0.03		
400	1000.89	0.46	1000.00	0.89	-3	-86.33	0.02		
500	1251.58	0.51	1250.00	1.58	-1	-28.77	0.02		
600	1502.16	0.50	1500.00	2.16	0	0.06	0.02		
700	1752.56	0.51	1750.00	2.56	1	28.83	0.02		
800	2003.37	0.50	2000.00	3.37	3	86.35	0.02		
900	2253.58	0.51	2250.00	3.58	5	143.88	0.02		
1000	2504.37	0.60	2500.00	4.37	10	287.76	0.02		
1100	2754.79	0.54	2750.00	4.79	15	431.54	0.02		
1200	3005.37	0.50	3000.00	5.37	PIR				
1300	3255.84	0.37	3250.00	5.84	-1	-213.66	0.09		
1400	3506.42	0.51	3500.00	6.42	-0.5	-109.36	0.14		
CHANNEL 6					0	-4.40	0.18		
0	0.00	0.00	0.00	0.00	0.5	100.02	0.13		
1	0.73	0.46	1.97	-1.24	1	204.53	0.11		
10	18.42	0.51	19.70	-1.28	2	413.42	0.17		
50	97.47	0.51	98.50	-1.03	2.4	496.75	0.15		
100	196.42	0.51	197.00	-0.58	CASE 499400 ohms reference				
250	492.84	0.50	492.50	0.34	6000	59.11	0.01		
500	987.32	0.48	985.00	2.32	8000	78.50	0.02		
750	1481.32	0.48	1477.50	3.82	10000	97.70	0.01		
1000	1975.79	0.42	1970.00	5.79	12000	116.80	.01		
1250	2469.84	0.37	2462.50	7.34	15000	145.13	.01		
1500	2964.16	0.60	2955.00	9.16	20000	191.69	.01		
1750	3458.58	0.51	3447.50	11.08	25000	237.27	.01		
2000	3952.79	0.42	3940.00	12.79	30000	281.96	.01		
CHANNEL 7					35000	325.80	0.01		
0	0.00	0.00	0.00	0.00	DOME 499400 ohms reference resistor				
1	1.44	0.51	2.96	-1.52	6000	59.13	0.01		
5	13.50	0.51	14.80	-1.30	8000	78.57	0.02		
10	28.10	0.45	29.60	-1.50	10000	97.84	0.01		
25	72.33	0.48	74.00	-1.67	12000	116.96	.01		
50	145.95	0.52	148.00	-2.05	15000	145.38	.01		
100	293.47	0.61	296.00	-2.53	20000	192.03	.01		
200	588.37	0.50	592.00	-3.63	25000	237.73	.01		
300	882.95	0.52	888.00	-5.05	30000	282.54	.01		
400	1177.68	0.58	1184.00	-6.32	35000	326.53	0.01		
500	1472.79	0.42	1480.00	-7.21	VREF RESISTORS - chan 6				
600	1767.40	0.60	1776.00	-8.60	99970	9993			
700	2062.16	0.37	2072.00	-9.84	END				
800	2356.89	0.46	2368.00	-11.11					
900	2651.79	0.54	2664.00	-12.21					