

CALIBRATION REPORT
Portable Radiation Package

SERIAL NUMBER 05

DataSet Configuration: 0101
Document date: February 8, 2001

Configuration File:

PRP CALIBRATION INFORMATION FILE -- created: 2001-02-07 (038) 13:25:02
LAST EDIT: new file
CALIBRATION NAME: 0101
DATA POINT OF CONTACT: R. M. Reynolds
POC ADDRESS: 490D; Brookhaven National lab; Upton NY 11973; USA
POC EMAIL: reynolds@bnl.gov
PRP SERIAL NUMBER: 05
FILE NAME: p05_calinfo_0101.txt
DOCUMENTS: Cal05_0101.pdf
HEAD SERIAL NUMBER: 461
HEAD CALIBRATION ID: 0101
DATALOGGER SERIAL NUMBER: dl00_6
DATALOGGER CALIBRATION ID: 0101
PSP SERIAL NUMBER: 31282F3
PSP CALIBRATION ID: orig
PIR SERIAL NUMBER: 30169F3
PIR THERMOPILE CALIBRATION ID: 0101
PIR TEMPERATURE CALIBRATION ID: 0101
COMMENTS: Test case



Bldg 490d, Upton NY 11973 — 631-344-7836 — reynolds@bnl.gov

PRP INFO FILE

SETUP FOR PROCESSING PRP CALIBRATIONS: 07-Feb-2001 16:18:53
PRP S/N: 05, Calibration identifier: 0101
Configuration file: hd:instruments:prp:prpcal:prp:05:0101:INFO_05_0101.txt
PRP CALIBRATION
PRP CALIBRATION INFORMATION FILE -- created: 2001-02-07 (038) 13:25:02
LAST EDIT: new file
CALIBRATION NAME: 0101
DATA POINT OF CONTACT: R. M. Reynolds
POC ADDRESS: 490D; Brookhaven National lab; Upton NY 11973; USA
POC EMAIL: reynolds@bnl.gov
PRP SERIAL NUMBER: 05
FILE NAME: p05_calinfo_0101.txt
DOCUMENTS: Cal05_0101.pdf
HEAD SERIAL NUMBER: 461
HEAD CALIBRATION ID: 0101
DATALOGGER SERIAL NUMBER: dl00_6
DATALOGGER CALIBRATION ID: 0101
PSP SERIAL NUMBER: 31282F3
PSP CALIBRATION ID: orig
PIR SERIAL NUMBER: 30169F3
PIR THERMOPILE CALIBRATION ID: 0101
PIR TEMPERATURE CALIBRATION ID: 0101
COMMENTS: Test case

CALIBRATION INFO FOR HEAD 461:

This file : 461.CAL
Data valid from date : 01/11/2001
MFRSR system owner : UNIV. OF MIAMI
YESDAS system password: Langley!
Supervisor password : Irradiance!
System Datalogger ID : \$0000 (Hex), 0 (Dec)
Instrument Head ID : \$0A70 (Hex), 2672 (Dec)
Instrument Head S/N : 461

DATALOGGER CALIBRATION: ProcLoggerCal (version 101) Run date: 07-Feb-2001 16:23:32
DATALOGGER S/N DL2000_6, PREAMP 6 (Minnett PRP05)
CALIBRATION DATE: 010122
TECHNICIAN: EDWARDS
VOLTAGE REFERENCE: VOLT-A-VIDER (cal: ???)\
USE PRECISION VREF CIRCUIT

CALIBRATION INFO FOR HEAD 461:

This file : 461.CAL
Data valid from date : 01/11/2001
MFRSR system owner : UNIV. OF MIAMI
YESDAS system password: Langley!
Supervisor password : Irradiance!
System Datalogger ID : \$0000 (Hex), 0 (Dec)
Instrument Head ID : \$0A70 (Hex), 2672 (Dec)
Instrument Head S/N : 461

```

-----
DATALOGGER CALIBRATION: ProcLoggerCal (version 101) Run date: 07-Feb-2001 16:23:32
DATALOGGER S/N DL2000_6, PREAMP 6 (Minnett PRP05)
  CALIBRATION DATE: 010122
  TECHNICIAN: EDWARDS
  VOLTAGE REFERENCE: VOLT-A-VIDER (cal: ???)\
  USE PRECISION VREF CIRCUIT

```

```

-----
PSP CALIBRATION: S/N 31282F3
  Factory cal: 2000 8.51, 2001-01-23 8.39
8.39
  8.390

```

```

-----
PIR CALIBRATION - S/N: 30169F3
  New calibration 2001-01-23 4.08
4.08
  4.080

```

```

-----
PIR THERMISTOR CALIBRATION -- PART 2

```

```

Process times from file lacie:instruments:prp:prpcal:pir:30169F3:30169F3_Tcal_0101.dat
SAVE THE RESULTS TO FILE lacie:instruments:prp:prpcal:pir:30169F3:30169F3_Tcal_0101.mat
save lacie:instruments:prp:prpcal:pir:30169F3:30169F3_Tcal_0101.mat readme beta_case beta_dome rcav rcs
  beta_case and beta_dome are Steinhart-Hart coefficients
  tcase = steinhart(hart(beta_case,r) where r = thermistor resistance in ohms.
  tdome = steinhart(hart(beta_dome ...
  rcav and rcstd are the test resistances and the std dev for each test point.
  rdav and rdstd are ditto for the dome
  tav and tstd are the mean temperature and std dev for each test point as
  computed from the SBE834 temperature probe.
  tc_ysi and td_ysi are computed from the published YSI calibration table.
  (See function 'YSI44006.m'.)
  casefit and domefit are cubic fitted corrections to the YSI computed temperatures.
  If tcseysi is a vector of case temperatures computed from the YSI44006.m function,
  then tcase = tcase + polyval(casefit,tcseysi);
  Ditto for the dome.

```

```

-----
COMPUTE PIR THERMISTOR COEFFICIENTS

```

```

Use calibrated coefficients for PIR case thermistor

```

```

Use calibrated coefficients for PIR doome thermistor

```

ZENITH ANGLE ERROR PLOTS

Head S/N: 461

Cal date: 01/11/2001

Now: 07-Feb-2001 15:10:58




Figure 1: Zenith Angle Error




Figure 2: Zenith Angle Error

HEAD 461 TOA IRRADIANCES BASED ON ASTRONOMICAL SOLAR SPECTRUM

WAVELENGTH (nm)			IRRADIANCE (W/m ² /nm)		
LOWER	CENTER	UPPER	LOWER	MEAN	UPPER
403,	414,	426,	1.652,	1.739,	1.826
484,	496,	508,	1.850,	1.947,	2.045
597,	614,	644,	1.627,	1.713,	1.798
649,	670,	830,	1.457,	1.534,	1.611
848,	868,	880,	0.907,	0.954,	1.002
926,	941,	955,	0.779,	0.820,	0.861







Figure 3: Head and Logger combined gains



$$1/(T+T_0) = p_1 a^3 + p_2 a^2 + p_3 a + p_4$$

$a = \ln(\text{mvadc}), T_0 = 273.15$

Case: max err = 1.807 C

$$p_1 = -2.4364e-04, p_2 = 3.8852e-03$$

$$p_3 = -2.0204e-02, p_4 = 3.7644e-02$$

Dome: max err = 1.807

$$p_1 = -2.4322e-04, p_2 = 3.8784e-03$$

$$p_3 = -2.0168e-02, p_4 = 3.7580e-02$$

Figure 4: Head and Logger combined gains





Figure 5: PIR Temperature calibration results.



TEST PLUG:

$R_{case} =$ _____ $T_{44031} =$ _____ Logger Temperature = _____

$R_{dome} =$ _____ $T_{44031} =$ _____ Logger Temperature = _____

COMPUTE VREF FROM R_REF AND R_THERM

PRP SN: 05 Cal ID: 0101

CASE - Rref = 499700.0

R_cal	v_T	V_REF (computed)
6000	109.3	9216.4
8000	144.7	9183.0
10000	179.5	9147.6
15000	213.8	7335.8
20000	264.3	6868.9
25000	346.3	7267.5
30000	425.3	7509.0

DOME - Rref = 499700.0

R_cal	v_T	V_REF (computed)
6000	109.3	9213.0
8000	144.6	9178.6
10000	179.4	9143.5
15000	213.7	7332.1
20000	264.2	6865.8
25000	346.1	7264.2
30000	425.1	7505.7

(File: hd:instruments:prp:prpcal:prp:05:0101:TcalVref_0101.dat)

```
% CALIBRATION FILE FOR PRPRX DATA COLLECTION SOFTWARE
% PSP CALIBRATION, PSP SN: 31282F3
4.1288      -1.3535
% PIR CALIBRATION, PIR SN: 30169F3
1.1728  5.1004
% TCASE FIT
-0.000243636    0.00388522    -0.0202043    0.0376439
% TDOME FIT
-0.000243219    0.00387844    -0.0201683    0.0375803
% K COEFFICIENT
4.0
% SIGMA
5.67e-8
% EPSILON
0.98
% BATTERY
0.030820 0.0
```

(File: hd:instruments:prp:prpcal:prp:05:0101:prprx_05_0101.txt)

LOGGER CALIBRATION FILE

```

% MFRSR: 00461                250  989.71 0.62
% PSP: 31282F3, COEFF: 8.39    300  1188.12 0.76
% PIR: 30169F3, COEFF: 4.08   350  1386.41 0.95
% Filename: DL2000_6.txt       400  1584.70 0.79
DATALOGGER S/N DL2000_6, PREAMP 6 (Minneapolis) 450 1702.93 1.13
  CALIBRATION DATE: 010122     500  1981.26 1.08
  TECHNICIAN: EDWARDS          600  2378.25 0.99
  VOLTAGE REFERENCE: VOLT-A-VIDER (cal: 7700) \ 700 2774.96 0.84
CHANNEL 1                      800  3171.78 0.81
0 0 0.0                        900  3568.65 0.80
10 46.05 1.02                  1000 3965.28 0.96
50 243.32 1.18                  CHANNEL 4
75 366.74 0.86                  0 0 0.0
100 489.59 1.40                 1  1.00 0.85
150 736.84 0.90                 5  13.27 1.12
200 983.10 1.08                 10 28.23 0.92
250 1229.94 1.16                25 72.17 1.14
300 1476.47 1.38                50 146.02 1.11
350 1723.34 1.03                75 220.02 1.16
400 1970.03 1.02                100 294.13 1.17
450 2216.94 1.03                200 589.67 1.20
500 2463.53 0.99                250 737.47 1.11
600 2956.83 1.02                400 1181.29 1.20
700 3450.21 0.96                650 1920.61 0.74
750 3696.42 1.18                900 2659.37 1.18
CHANNEL 2                      1000 2954.87 1.31
0 0 0.0                        1100 3251.00 1.14
10 96.12 1.21                   1200 3546.37 1.25
50 493.60 0.79                   1300 3841.94 1.26
75 741.48 0.89                  CHANNEL 5
100 989.61 1.20                  0 0 0.0
150 1486.00 0.91                 1  0.86 0.86
200 1981.87 1.24                 5  10.48 1.27
250 2478.55 1.20                 10 22.96 1.16
300 2975.12 0.91                 25 60.96 1.17
350 3470.84 1.14                 50 123.56 1.31
400 3967.69 1.03                 100 248.53 1.16
CHANNEL 3                      200 499.97 1.01
0 0 0.0                        400 1002.28 1.02
10 37.47 0.73                   500 1252.80 1.02
25 96.89 1.16                   750 1880.69 1.18
50 195.98 1.01                  1000 2508.50 1.21
75 295.42 0.88                  1100 2759.49 1.13
100 394.58 0.78                 1200 3010.69 0.96
150 592.89 0.84                 1300 3261.52 1.10
200 791.05 0.92                 1400 3512.68 1.05

```

1500	3763.78	1.03	-1	-28.60	0.02
CHANNEL 6			0	0.39	0.02
0	0	0.0	1	29.24	0.02
1	0.20	0.4	3	86.95	0.02
5	7.38	1.09	5	144.68	0.02
10	17.12	1.20	10	289.03	0.02
25	47.10	0.98	15	433.30	0.02
50	96.48	1.09	PIR		
75	145.93	1.20	-1	-213.52	0.11
100	195.28	1.02	-.5	-109.25	0.09
250	491.97	1.20	0	-3.82	0.12
500	986.79	1.04	0.5	100.25	0.13
750	1481.71	1.21	1.0	204.85	0.14
1000	1976.15	0.95	2.0	413.70	0.10
1200	2372.04	0.81	2.4	496.92	0.09
1600	3163.31	0.84	CASE 499700 ohms reference		
1800	3559.00	0.93	6000	109.35	0.02
2000	3954.79	0.73	8000	144.7	0.01
CHANNEL 7			10000	179.47	0.02
0	0	0.0	15000	213.79	0.01
1	1.08	0.83	20000	264.34	0.00
5	12.96	1.04	25000	346.27	0.01
10	28.00	1.06	30000	425.28	0.00
25	72.34	1.07	DOME 499700 ohms reference resistor		
50	146.81	0.76	6000	109.31	0.01
75	220.75	1.11	8000	144.63	0.01
100	294.77	1.03	10000	179.39	0.02
250	739.00	1.11	15000	213.68	0.01
500	1480.16	0.85	20000	264.22	0.01
750	2221.17	0.80	25000	346.11	0.01
1000	2961.65	0.98	30000	425.09	0.02
1200	3554.37	0.90	VREF RESISTORS - chan 6		
1300	3850.89	0.82	99960	9993	
PSP			END		
-5	-144.07	0.03			
-3	-86.27	0.11			