

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ac-s Calibration

Date	3/16/2016	Customer	University of California Santa Barbara	RO #	029420
S/N #	ACS-001	Technician	Cad/JGM	Cal #	7

Wavelength range:	400 to 730 nm	
Tcal:	16.1 deg C	
Ical:	24.4 deg C	
Acal:	27.7 deg C	
Temperature range:	7.6 to	35.0 deg C
Zero:	25.5 deg C	

Definitions

Wavelength range: Wavelengths for which instrument use is recommended.

Tcal: Water temperature in flow tubes during water calibration.

Ical: Internal instrument temperature during water calibration.

Acal: Internal instrument temperature when recording air offsets in the *.cal file.

Temperature range: Recommended internal instrument operating range.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.